

PBL Pacific BioLabs

The Service Leader in Bioscience Testing

SPONSOR 13609

Soong Soi Peng
LOTTE CHEMICAL TITAN (M) SDN.BHD.
PLO8, Tanjung Langsat Industrial Estate
Pasir Gudang, Johor, 81700
Malaysia

REPORT DATE: 7/8/14



MSP NUMBER: 14E0269H

P.O. NUMBER: 4500472576

PAGE: 1 of 1

CERTIFICATE

CLASS VI 70°C PLASTIC TESTS

(USP 37 <88>, REV. 5/2014)

Name: TITANPRO PD701
Physical Description: Solid
Total Quantity Received for Testing: 1 bag containing 100 g
Lot Number: A140408A
Storage Condition: Room Temperature
Date Received: 5/21/14

Class VI – Intracutaneous Test

PBL Report No.: 14E0269H-X02G

Class VI – Systemic Injection Test

PBL Report No.: 14E0269H-X03G

Class VI – Intramuscular Implantation Test

PBL Report No.: 14E0269H-X04G

The test article met the requirements for Plastic Class VI - 70°C testing per USP <88>.



Multiple Study Project Coordinator
Pacific BioLabs



Date



QAU Review



Date

:sdb